



PullTester 326

Pull Testing Machine

- Simple LCD display for easy programming and digital pull force read out
- Speed-controlled motor for consistent pull rates throughout the measuring range
- Triple range for improved accuracy over a wider range of wires
- RS 232 interface for curve analysis and statistics with WinCrimp software
- 4 selectable pulling rates depending on test requirements
- 4 pulling modes for destructive and non-destructive tests
- Memory for up to 2400 values
- Networking capabilities

QUALITY ASSURANCE

PullTester 326

Concept

Schleuniger's PullTester 326 is a triple-range, motorised, bench-top unit designed to measure pull-test forces of crimp and ultrasonic weld connections on a wider range of wires than single-range pull test devices. Pull test values are critical parameters for quality control and assurance. The PullTester 326 can also perform non-destructive tests (hold to a specified force). This versatile machine has three measuring ranges, which are individually calibrated. The standard measuring ranges are 200N, 1000N, and 2000N (45lb, 220lb, 440lb) however other ranges are possible. This triple-range capability ensures the highest accuracy for the widest range of applications. Hand actuated or pneumatic pull test devices can give inconsistent data depending on the operator or pull rate. Some standards specify that a test device must pull with a consistent rate. The Schleuniger PullTester 326 is equipped with a speed-controlled motor, ensuring consistent pull rates throughout the measuring range resulting in repeatable and accurate data. Pull forces can be measured in pounds, Newtons or kiloponds. The standard 14-position terminal holder accommodates a wide variety of terminals to suit most applications. A variety of terminal holders, however, are available upon request.

Application

The PullTester 326 has features such as four pulling rates and internal memory to accommodate more stringent test requirements. It can also be integrated with a quality network which brings together crimp height, pull test and crimp force data to ensure a high quality tested product. Pull test data can be stored for future reference or downloaded for statistical evaluation. The PullTester 326 will test pull forces up to 2000 Newtons (440 lbs.) and is specially suited for quality assurance in a production environment.

Technical specifications	
Measuring Range	Standard: 0 – 200 N, 0 – 1000 N, 0 – 2000 N (0 – 45 lbs., 0 – 220 lbs., 0 – 440 lbs.) other variations possible by special order
Unit of Measure	Newton (N), Kilopond (Kp), Poundforce (lbf.)
Display	Upper: LCD 6-digit for force readings Lower: LCD 4-line for programming and operation
Pressure Precision	0.5% of measurement range
Operating Temperature	0 – 50°C
Stroke	43 mm (1.69")
Pulling Speed	4 speeds: 25, 50, 100 mm/min. or high speed. (1.97, 2.95, 3.94"/min. or high speed)
Pulling Modes	Pull + Break: Normal pull test until wire breaks Pull + Hold: Pull to a specified force and hold for 1 sec - 252 min (non-destructive test) Pull + Return: Pull to a specified force and reduce (non-destructive test) Pull + Hold + Break: Pull to a specified force and hold for 1 sec - 252 min and pull until wire breaks
Memory Capacity	up to 48 jobs with 50 measurements (2400 measurements)
Monitoring	Device display output; Optional WinCrimp statistical software for visual force-time-table on PC and statistical analysis for evaluation with download possibility to Microsoft® Excel software.
Write Protection	IP 20
Print Capability	RS 232 connection directly to printer or via PC using WinCrimp Software
Network	Multiple devices in combination with crimp force monitor and crimp-height measurements device via WinCrimp software with either RS232 or TCP/IP.
Interfaces	RS 232
Motor	24 VDC
Weight	Approximately 11 kg. (24 lbs.)
Dimensions	250 x 130 x 410 mm (10 x 5 x 16")
CE-Conformity	The PullTester 326 fully complies with all CE and EMC equipment guidelines relative to mechanical and electrical safety and electromagnetic compatibility.
Important Note	Schleuniger recommends that wire samples be submitted in cases where there is doubt as to the processing capabilities of a particular machine.